CALIBRATION WORK SHEET

Date of Calibration: 9/13/Ders	Sonde ID: \$368061#4
Technician: Gelp	
7777	3 to 6 hours before calibrating for unattended
	s; run in Discrete mode for 10 minutes to accelerate pid Pulse DO Only)
Turbidity wiper changed? Y N Chlorophyll wipe	er changed? Y N
BGA-PC wiper changed? Y N Rhodamine wiper	r changed? Y N
Note: If parking problems occur with optical probes having a serial number 3.06 or later. Parking issues with optical probes having a serial number pad.	umber 07L (Dec 07) or older, be sure the firmware is er prior to 07L may be related to a dirty wiper body or
Record sonde battery voltage:// Vc (if appli	Standard Pre Cal / Post Cal
Record the following diagnostic numbers after calibrat 6560 Conductivity cell constant S, O 7216 Range 5.0 ± .45	TemperatureSonde
Integrated conductivity cell constant Range $5.0 \pm .70$	Conductivity 50 48,76 50.00
pH mv Buffer 7 -39 Range 0 ± 50 mv	pH 7 7 7.5/26
pH-mv Buffer 4 Range +180 ±50 mv*	pH 4/
pH mv Buffer 10/88.5 Range -180 * 50 mv *	pH 10 10 9.65/ 10
*Note: Millivolt span between pH 4 and 7 should be \approx 165 to 180 mv	ORP /
Millivolt span between pH 7 and 10 should be ≈ 165 to 180 m	v Turbidity /
DO charge (RP only) Range 25 to 75	Turbidity /
DO gain 1.4 93 Range 0.7 to 1.4	Turbidity 0.5
ODO gain Range 0.85 to 1.15	Chlorophyll /
	Chlorophyll
Turbidity standard used in calibration	DORP Sut 13821 99.69
Manufacturer and part number	DO ROX
	BGA PE/PC /
Barometric Pressure: 755.65 mmHg	BGA PE/PC
DO % Calculated – (BARO mmHg divided by 7.6) = % saturation Example: $760 \div 7.6 = 100.0\%$	Rhodamine/
Depth Calibration - If zero was entered, record barometric pressure at t	time of calibration mmHg
Depth Calibration - If offset depth was entered, record value	
Depth Calibration (Vented) - Acceptable calibration constant: 0.0 ps	
Notes: to a vertain to many	Elea 1 deell
Notes: timing for YSI was 6 minutes Ogain is out of Range pH offset	- The same of the
Ogain is out of Kange pH offset	(pH-7)*/x = -22345/