



Calibration Worksheet



pH 20A102061

	UTC Time	Coordinated Universal Time
Start Date/Time	6/26/2020 09:20:18	6/26/2020 09:20:18
End Date/Time	6/26/2020 09:27:44	6/26/2020 09:27:44
Previous Calibration Date/Time	6/20/2020 07:00:51	6/20/2020 07:00:51

Sensor Type	pH
Sensor SN	20A102061
Sensor Firmware Version	3.0.0
Calibration Parameter	pH

Sonde Type	EXO2 Sonde
Sonde SN	17D108371
Sonde Firmware Version	1.0.83
Sonde ID	Sonde 17D108371 CM



QC Score

	Cal Point 1	Cal Point 2	Cal Point 3
Standard	7.00 pH	10.01 pH	
Pre Calibration Value	7.05 pH	10.07 pH	
Post Calibration Value	7.00 pH	10.01 pH	
Raw Value (pH mV)	-11.42	-188.68	
Temperature	23.93 ° C	23.86 ° C	
Additional Input 1 (N/A)			
Additional Input 2 (N/A)			
Additional Input 3 (N/A)			
Type	YSI pH 7 Buffer	YSI pH 10 Buffer	
Manufacturer	YSI	YSI	
Lot Number	20D3S	20D4T	
Calibration Point Accepted	YES	YES	
Stability Achieved	YES	YES	

Completed	YES
Applied	YES
Valid	YES
Sensor Removed	NO
Uncalibrated	NO

Additional Post Calibration Info:
 pH 7.00 and 10.01 Delta Slope 177.25 mV
 58.89 mV per pH unit

Hardware	EXO Desktop
KOR Version	1.0.12
Worksheet Version	1



Calibration Worksheet



pH 17F100785

	UTC Time	Coordinated Universal Time
Start Date/Time	6/26/2020 09:20:19	6/26/2020 09:20:19
End Date/Time	6/26/2020 09:29:43	6/26/2020 09:29:43
Previous Calibration Date/Time	6/20/2020 07:02:12	6/20/2020 07:02:12

Sensor Type	pH
Sensor SN	17F100785
Sensor Firmware Version	3.0.0
Calibration Parameter	pH

Sonde Type	EXO2 Sonde
Sonde SN	17D108371
Sonde Firmware Version	1.0.83
Sonde ID	Sonde 17D108371 CM



QC Score

	Cal Point 1	Cal Point 2	Cal Point 3
Standard	7.00 pH	10.01 pH	
Pre Calibration Value	7.05 pH	10.07 pH	
Post Calibration Value	7.00 pH	10.01 pH	
Raw Value (pH mV)	-5.95	-183.14	
Temperature	23.93 ° C	23.86 ° C	
Additional Input 1 (N/A)			
Additional Input 2 (N/A)			
Additional Input 3 (N/A)			
Type	YSI pH 7 Buffer	YSI pH 10 Buffer	
Manufacturer	YSI	YSI	
Lot Number	20D3S	20D4T	
Calibration Point Accepted	YES	YES	
Stability Achieved	YES	YES	

Completed	YES
Applied	YES
Valid	YES
Sensor Removed	NO
Uncalibrated	NO

Hardware	EXO Desktop
KOR Version	1.0.12
Worksheet Version	1

Additional Post Calibration Info:
 pH 7.00 and 10.01 Delta Slope 177.19 mV
 58.87 mV per pH unit



Calibration Worksheet



pH 18M102562

	UTC Time	Coordinated Universal Time
Start Date/Time	6/26/2020 09:20:18	6/26/2020 09:20:18
End Date/Time	6/26/2020 09:28:54	6/26/2020 09:28:54
Previous Calibration Date/Time	6/20/2020 07:01:31	6/20/2020 07:01:31

Sensor Type	pH
Sensor SN	18M102562
Sensor Firmware Version	3.0.0
Calibration Parameter	pH

Sonde Type	EXO2 Sonde
Sonde SN	17D108371
Sonde Firmware Version	1.0.83
Sonde ID	Sonde 17D108371 CM



QC Score

	Cal Point 1	Cal Point 2	Cal Point 3
Standard	7.00 pH	10.01 pH	
Pre Calibration Value	7.30 pH	10.21 pH	
Post Calibration Value	7.00 pH	10.01 pH	
Raw Value (pH mV)	64.10	228.87	